



DQS Medizinprodukte GmbH | August-Schanz-Str. 21 | 60433 Frankfurt am Main

## Boston Medical Products, Inc.

70 Chestnut Street  
Shrewsbury, MA 01545  
United States of America

Date: 2024.05.07

### Notified Body Confirmation Letter

**Reference: 1000177467**

To whom it may concern,

**Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices**

This letter confirms that, DQS Medizinprodukte GmbH, a Notified Body designated against Regulation (EU) 2017/745 (MDR) and identified by the number 0297 on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

## Boston Medical Products, Inc.

70 Chestnut Street  
Shrewsbury, MA 01545  
United States of America

SRN: US-MF-000017994

The devices covered by the formal application and the written agreement mentioned above are identified in the Tables listed below: Table 1 identifies the devices for which an MDR application has been received, written agreement concluded and for which DQS Medizinprodukte GmbH is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive. Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but DQS Medizinprodukte GmbH has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive.

In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired after 26 May 2021 and before 20 March 2023, without having been withdrawn, this letter also confirms that the manufacturer signed the written agreement under MDR by the date of MDD/AIMDD certificate expiry, or provided evidence that a competent authority of a Member State had granted a derogation or exemption from the applicable conformity assessment procedure in accordance with Article 59(1) of MDR or Article 97(1) of the MDR respectively, by the 20 Mar 2023 for the relevant devices. The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3c of MDR (as amended by (EU) 2023/607), are shown below:

- 26 May 2026 for Class III custom-made implantable devices

- 31 December 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31 December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- 31 December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of the Notified Body,

A handwritten signature in black ink, appearing to read 'M. Köhler'.

**Melina Köhler**

Regulatory Affairs Manager

**Table 1: Devices covered by this letter and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:**

<b>Device name and Basic UDI-DI (as proposed by the manufacturer within the application)</b>	<b>MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)</b>	<b>If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device</b>	<b>MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification</b>
Epistaxis Catheter 4063107EPC8H	Class I devices placed on the market in sterile condition	Rhinology Catheters	488278 MR2, ID#170773614, NB 0297
Intranasal Splints 4063107NAS9H	Class IIa	Nasal Splints	488278 MR2, ID#170773614, NB 0297
Montgomery® Long-Term Cannulas 4063107TCSAM	Class IIb excluding Class IIb implantable non-WET	Tracheostomy Tubes, Cannulas Tracheostomy Speaking Valve	488278 MR2, ID#170773614, NB 0297
Moore Tracheostomy Tubes 4063107MTTB7	Class IIb excluding Class IIb implantable non-WET (not verified yet)	Tracheostomy Tubes	488278 MR2, ID#170773614, NB 0297
Montgomery® Salivary Bypass Har-EL pharyngeal Tube 4063107EST9U	Class IIb implantable non- WET device (not verified yet)	Esophageal Tubes	488278 MR2, ID#170773614, NB 0297
Montgomery® Safe-T-Tubes™ 4063107STTC5	Class IIb implantable non- WET device (not verified yet)	Tracheostomy Tubes	488278 MR2, ID#170773614, NB 0297
Montgomery® Thyroplasty Implant 4063107THYBG	Class IIb implantable non- WET device (not verified yet)	Thyroplasty Implants Thyroplasty Instruments	488278 MR2, ID#170773614, NB 0297
Montgomery® Thyroplasty Measuring Kit 4063107THYBG	Class IIa	Thyroplasty Implants Thyroplasty Instruments	488278 MR2, ID#170773614, NB 0297
Montgomery® Thyroplasty Instrument Set 4063107THYBG	Class I devices that qualify as re-usable surgical instruments	Thyroplasty Implants Thyroplasty Instruments	488278 MR2, ID#170773614, NB 0297

**Table 2: Devices covered by this letter and for which the NB is NOT responsible for appropriate surveillance of the corresponding devices under the applicable Directive:**

Device name and Basic UDI-DI (as proposed by the manufacturer within the application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
N/A	N/A	N/A	N/A

**Confirmation Letter Revision History**

Date	NB internal reference traceable to each version of the letter	Action
2023-07-13	1000129013	Initial issue
2024-05-07	1000177467	addition of device